

## **Declaration of Conformity**

---

**Biotage Sweden AB, Box 8, SE-751 03 Uppsala, Sweden**

declare under our sole responsibility that the product

**Biotage® PhyPrep™**

**P/N: 416700**

to which this declaration relates are in conformity with the following directives, standards and other normative documents;

**Directive 2006/42/EC**

Standards used:

**Machinery Directive (MD)**

EN 61010-1:2010/A1:2019  
EN 61010-2-081:2020

**Directive 2014/30/EU**

Standards used:

**Electro Magnetic Compatibility (EMC)**

EN 61326-1:2013  
CFR 47 FCC part 15 Subpart B. Class A

**Directive 2012/19/EU**

Standard used:

**WEEE**

EN 50419:2006

**Directive 2011/65/EU +  
2015/863/EU**

Standard used:

**RoHS**

EN 63000:2018

**For Canada and the USA**

UL 61010-1:2012 /R:2019-07  
UL 61010-2-081:2019  
IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016  
IEC 61010-2-081: 2019

CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11  
CAN/CSA-C22.2 No. 61010-2-081:2019

CFR 47 FCC part 15 Subpart B. Class A

**NTRL Certificate No. U8 071091 0037 Rev. 00 issued by TÜV SÜD Product Service GmbH**

**CB Certificate No. DE 3 – 31447 issued by TÜV SÜD Product Service GmbH**

**Uppsala, 26 February 2021**

  
.....  
(Jon-Sverre Schanche, Chief Scientific Officer - Biotage)