

## Declaration of Conformity

---

**Biotage Sweden AB, Box 8, SE-751 03 Uppsala, Sweden**

declare under our sole responsibility that the product

<b>Biotage® Initiator+</b>	<b>356700</b>
<b>Biotage® Initiator+ Alstra</b>	<b>356701</b>
<b>Biotage® Robot 8</b>	<b>355380</b>
<b>Biotage® Robot 60</b>	<b>355381</b>
<b>Biotage® Alstra Robot</b>	<b>356356</b>
<b>Biotage® SP Wave</b>	<b>356013</b>

to which this declaration relates are in conformity with the following directives, standards and other normative documents:

**Directive 2014/35/EU**

Standards used:

**Low Voltage Directive (LVD)**

EN 61010-1:2010  
EN 61010-2-010:2014  
EN 61010-2-051:2015  
EN 61010-2-081:2015

**Directive 2014/30/EU**

Standards used:

**Electro Magnetic Compatibility (EMC)**

EN/IEC 61326-1:2013, Class B  
CFR 47 FCC Part 18, Subpart C

**Directive 2012/19/EU**

Standard used:

**WEEE**

EN 50419:2006

**Directive 2011/65/EU +  
2015/863/EU**

Standard used:

**RoHS**

EN 63000:2018  
GB/T 26572-2011 (China RoHS)

**For Canada and the USA**

UL 61010-1:2012-05  
CAN/CSA-C22.2 No. 61010-1:2012-05  
IEC 61010-1:2010 (CB)  
IEC 61010-2-010:2014 (CB)  
IEC 61010-2-051:2015 (CB)  
IEC 61010-2-081:2015 (CB)

CFR 47 FCC part 15 Subpart B. Class A

Including national differences for Switzerland, Japan, Austria, Denmark, Republic of Korea, Slovenia, Sweden, United Kingdom

**UL listing and cUL listing report no: E491094-D1000-1/A1/C3-ULCB  
CB Certificate No. DK-63486-M1-UL issued by UL 2019-10-16**

**Uppsala, 06 November 2020**



(Jon-Sverre Schanche, Chief Scientific Officer - Biotage)